



Characterisation of conductive polymer devices using XPS

Nunney, Tim S.; White, Richard G.; Hansen, Thomas Steen; Larsen, Niels Bent

Publication date:
2007

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Nunney, T. S., White, R. G., Hansen, T. S., & Larsen, N. B. (2007). *Characterisation of conductive polymer devices using XPS*. Abstract from ECASIA '07, Brussels, Belgium.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.



Tuesday, September 11, 2007 – 09:00 – Room 3

TEC-1295

Characterisation of conductive polymer devices using XPS

Tim S Nunney¹, Richard G White¹, Thomas Steen Hansen^{2,3} & Niels B Larsen²

(1) Thermo Fisher Scientific, The Birches, Imberhorne Lane, East Grinstead, West Sussex, RH19 1UB, UK. (2) Danish Polymer Centre, Risoe National Laboratory, Technical University of Denmark, POL-313, P.O. Box 49, DK-4000 Roskilde, Denmark. (3) Department of Technical Engineering, Technical University of Denmark, DK-2800 Lyngby, Denmark - e-mail : tim.nunney@thermofisher.com

Conductive polymers may form the basis of new generations of active devices. The main advantages include low cost, ease of processing, and direct integration with many other polymer based functional materials. We have demonstrated a simple and versatile new method for fast micropatterning of conductive polymers. Features sizes down to 2 μm have been realised in a conductive polymer film (PEDOT) using a stamp impregnated with an aqueous chemical deactivation agent. The stamp is moulded in bas-relief which enables spatially selective transfer of the deactivation agent from the stamp to the conductive polymer in areas of contact. Conductive polymer films of thicknesses 250-750 nm were micropatterned with some variation in the resulting edge resolution. Identification and spatial mapping of the chemical products resulting from the deactivation procedure are crucial for further optimisation of the patterning process for applications in low-cost polymer electronic devices.

Surface analysis by XPS allows for imaging of surface features by chemical state, allowing the distribution of such chemical products to be mapped across the surface of the device. We present the results of new data processing algorithms which allow high spatial resolution data to be acquired whilst using a probe size much larger than the feature under investigation. Data can also be collected by operating at high pass energy to maximise the signal-to-noise ratio and using further algorithms to decouple the analyser contribution, resulting in spectra with high energy resolution. These advances allow for complex structures with features of a few microns in size to be imaged with significantly reduced acquisition times and much improved chemical state separation.